



Automated control of inspired oxygen (FiO₂) in preterm infants

Author: Carlo Dani^{1,2}

1. Division of Neonatology, Careggi University Hospital of Florence, Florence, Italy;
2. Department of Neurosciences, Psychology, Drug Research and Child Health, Careggi University Hospital of Florence, Florence, Italy.

Background

Oxygen is likely the most used medication in preterm infants.¹ For these patients, oxygen-therapy is a life-saving drug and at the same time a risk-factor for the development of severe complications such as bronchopulmonary dysplasia (BPD)² and retinopathy of prematurity (ROP).³ Therefore, the continuous monitoring of supplemental oxygen is very important; in preterm infants it is performed measuring the oxygen pulse saturation (SpO₂). Recently, the American Academy of Pediatrics suggested that a 90-95% SpO₂ represents a reasonable target⁴ in substantial agreement with European guidelines that suggest a 90-94% SpO₂ target.⁵ In fact, low (85-89%) target increases mortality and the risk of necrotizing enterocolitis (NEC), while a high target (91-95%) increases the risk of ROP.⁶

However, it is not easy to maintain the optimal SpO₂ target. In fact, the most important randomized controlled trials on this issue⁷⁻¹⁰ reported that patients spend a high percentage of time outside the planned target, ranging from 9.17 to 27.4%⁹ with SpO₂<85%, and from 11.1¹⁰ to 33%¹¹ with SpO₂ >95%. This occurs due to the respiratory instability of very preterm infants which causes SpO₂ fluctuations and the ineffective manual adjustments of FiO₂ by caregivers which can be late and defective, or prolonged and excessive. The consequence is patient exposure to recurrent

episodes of hypoxia (and associated bradycardia) and hyperoxia that can be detrimental for their outcome.^{2,3,11}

In fact, premature infants present with frequent episodes of spontaneous intermittent hypoxemia (SIH) which are more prevalent in infants who require prolonged mechanical ventilation and increase their frequency with postnatal age and with progressing chronic lung disease. A common mechanism triggering intermittent hypoxemia (SIH) is contractions of the abdominal muscles that opposes the respiratory activity of patients resulting in decreased lung volume, impaired lung mechanics, and hypoventilation. It has been observed that limited staff availability can further compromise SpO₂ targeting during routine clinical care enhancing the effect of SIH and that education of the clinical staff induces a relevant increase of time spent within the SpO₂ target range, decreases episodes of hyperoxia but, paradoxically, increases episodes of hypoxia.

The management of SIH include ventilator strategies and supplemental oxygen therapy. To increase high positive end-expiratory pressure (PEEP) allows to increase lung volume and basal oxygenation levels but is only partially effective in preventing SIH caused by active exhalations. In fact, the rise in intrathoracic pressure during forceful contraction of the abdominal muscles exceeds the level of PEEP considerably. High peak inspiratory pressure (PIP) or ventilator frequency

can limit SIH, but this strategy might provide excessive support during periods when the infant is not presenting with episodic hypoxemia and induce lung injuries.¹² Volume-targeted ventilation, which automatically adjusts the ventilator PIP to maintain a set tidal volume, can contribute to attenuate SIH compared with conventional pressure-controlled (PC) ventilation but is ineffective in preventing them. However, this required setting a target tidal volume (TV) that was larger than the TV delivered during mechanical ventilation, possibly exposing the infants to excessive TV and PIP during periods when ventilation is stable. High basal levels of SpO₂ can attenuate the frequency or severity of SIH, but this increases the exposure to high FiO₂ and hyperoxemia. It usually occurs that caregivers increase FiO₂ when SpO₂ decreases and, subsequently, maintain high FiO₂ to prevent SIH, but this strategy is ineffective in the medium- and long-term and SIH reappears soon.

Systems allowing the automated control of FiO₂

The aforementioned considerations represent the basis for development of systems allowing the automated control of FiO₂ patient delivery to maintain SpO₂ within a target range limiting episodes of both hypoxia and hyperoxia and decreasing the workload of nurses. These systems include a pulse oximeter, a non-invasive or invasive respiratory support (i.e. mechanical ventilator), and the algorithm that regulates the FiO₂ adjustments in relationship to SpO₂ changes.

1. Fathabadi et al. detailed the main characteristics of algorithms developed for automation of inspired oxygen control in preterm infants and classified them in four categories: rule-based (fuzzy or non-fuzzy algorithms), proportional-integral-derivative (PID), adaptive, and robust (this algorithm is not available in the clinical practice)¹³. Among the most diffuses systems, the Predictive Intelligent Control of Oxygenation (PRICO[®]) uses a rule-based non-fuzzy control algorithm (see figure 1) ;
2. OxyGenie[®] system use a PID algorithm; and the AVEA-CLiO2[®] uses an adaptive algorithm.^{14,15}




Figure 1 Flow scheme of PRICO algorithm, Hütten et al.¹⁵

Neonatal ventilators	Company	Algorithm
Fabian HFO	Acutronic, Hirzel, Switzerland	PRICO [®]
SLE6000	SLE Limited, South Croydon, UK	OxyGenie [®]
Avea	Vyaire Medical, CA, USA	CLiO2 [®]
Sophie	Fritz Stephan GmbH, Gackebach Germany	SPO2C [®]
Leoni plus	Heinen & Loewenstein GmbH, Bad Ems, Germany	CLAC [®]

Table 1 Largely diffused neonatal ventilators and integrated algorithms for automated control of fiO₂ delivery.

Thus, it is noteworthy to underline that different clinical effects of systems for automated control of FiO₂ in preterm infants may depend on many variables, such as study population, type of respiratory support, and chosen target range, but also reflect the effectiveness of the algorithm itself.

Clinical studies

Most studies compared the effectiveness of automated control for FiO₂ delivery versus manual control. Although they are heterogeneous for design, population size, duration, and device used, demonstrate that automated systems are significantly more effective than manual control in maintaining target SpO₂.¹⁴ Automated control has also been associated with less time spent with SpO₂ higher than the target in comparison with manual control. See Figure 2 below as an example Dani et al.¹⁸




Figure 2 Comparison of the SpO₂ outcome measurement during automated (with PRICO) (A) and manual (M) control of FiO₂. Time spent with SpO₂ 90%–95% was higher during the automatic than manual control of FiO₂, while time spent with SpO₂ <80% or >95% was lower.

This advantage has not always been observed for the time spent with SpO₂ lower than the target. Thus, it is possible that automated control is more effective in preventing hyperoxia than hypoxia in comparison with manual control.¹⁴ To explain why automated systems might perform less favourably with hypoxia, we must remember that closed loop systems can respond to episodes of oxygen desaturation but are ineffective in preventing them. Moreover, it is possible that the automatic increase of FiO₂ can attenuate the severity of hypoxemia but it is unlikely that it produces an immediate recovery of the episode. However, very recent studies by Dijckman et al.¹⁶ and Dani et al.¹⁷ found that PRICO[®] system was more effective of manual control in maintaining SpO₂ within target

range, but also decreased the duration of hypoxemia in preterm infants with frequent desaturations suggesting that some algorithms might be more effective of others in correcting low values of SpO₂.

It is interesting wondering to evaluate if the width of the SpO₂ target range (i.e.: > or > 6%) can affect the effectiveness of the automated system for FiO₂ control but available data do not seem to support this hypothesis.¹⁴ However, it seems reasonable to avoid too much wide target range which might mask worrying fluctuations in patients' oxygen need.

On the other hand, it would be interesting to compare the effectiveness of various algorithms but to date there are not many published studies on this topic. So far only Salverda et al. compared two different systems for automated control of FiO₂ delivery and found significant difference between them¹⁸ confirming that different algorithms can be associated with different results.

Clinical studies on effects of automated control of FiO₂ on tissue oxygenation

Some researchers investigated if a better control of SpO₂ results in an improvement of tissue oxygenation. Waitz et al. studied 15 very low birth weight infants who were randomly assigned to 24 h of automated (AVEA-CLiO2[®], Vyair medical, Yorba Linda CA, USA) or manual adjustment of FiO₂.¹⁹ They found that automated and manual FiO₂ control were associated to similar values and fluctuations of cerebral oxygenation.¹⁹ Gajdos et al. studied 12 very low birth weight infants who were randomly assigned to 24 h of automated (SPO2C[®], Sophie Respirator, Fritz Stephan GmbH Medizintechnik GmbH, Gackenbach, Germany) or manual adjustment of FiO₂.²⁰ They confirmed that automated and manual FiO₂ control were associated to similar values and fluctuations of cerebral, renal and hepatic rSO₂.²⁰ More recently, we studied 20 infants with gestational age <32 weeks who were randomly assigned to 24 h of automated (PRICO[®], fabian HFO[®], Acutronic Medical System AG, Hirzel, Switzerland) or manual adjustment of FiO₂ and, in agreement with previous studies, we found that cerebral and splanchnic oxygenation were similar during automatic and manual control of FiO₂.²¹

To explain why the more stable SpO₂ guaranteed by automated system of FiO₂ delivery did not improve cerebral oxygenation, we can speculate that endogenous mechanisms of cerebral compensation and autoregulation, which concur to maintain adequate cerebral blood perfusion and oxygenation despite abnormal arterial oxygenation, are per se very effective. Therefore, it is difficult that an automated FiO₂ control is able to further improve the stability of cerebral oxygenation. It is more difficult to explain the lack of effects of automated FiO₂ control on oxygenation of hepatic, renal and splanchnic tissues, since these district does not benefit from an autoregulation system that safeguards its blood perfusion. However, some explanations can be speculated: during the studies patients were in stable clinical condition and this might have helped limit fluctuations of tissue oxygenation;¹⁹⁻²¹ baseline values of hepatic, renal and splanchnic oxygenation are lower than those of cerebral oxygenation and, therefore, it is possible that significant changes can occur only for deep and prolonged changes of SpO₂ that both automated and manual control of FiO₂ can prevent.

Clinical studies on effects of automated control of FiO₂ on outcome of preterm infants.

No large studies investigated whether automated control of FiO₂ can affect outcome in preterm infants, and recently it has been questioned the actual possibility of these devices of beneficially influencing their neurodevelopment.²² In fact, it has been demonstrated that the correlation between SIH and neurodevelopment in preterm infants becomes significant at 6-10 weeks of postnatal age when oxygen supplementation and respiratory support are often absent or mild²³ and, consequently, the role of automated control of FiO₂ might be smaller than expected.

However, Salverda et al. have recently reported in a retrospective study that the implementation in their center of the automated oxygen control (CLiO₂ and OxyGenie® systems) did not change mortality and morbidity of patients but decreased the duration of mechanical ventilation.²⁴ This result is promising because prolonged mechanical ventilation is a risk factor for severe complications, such as bronchopulmonary dysplasia (BPD), and has been directly associated with poor neurodevelopmental outcome.

Conclusions

Automated devices are significantly more effective than manual control in maintaining SpO₂ within selected ranges and in preventing hyperoxia and to lesser extent hypoxia, although they do not improve tissue oxygenation. However, automated control can safely decrease the workload of nurses. Large studies are required to assess the actual effects of automated control of FiO₂ on infants' outcome. These RCTs should assess the effect of automated control of FiO₂ on short and long-term outcome in preterm infants including neurodevelopment at 2 years of life and later. Moreover, further clinical studies comparing the effectiveness of various algorithms would be more useful to assess their most effective settings.

REFERENCES

1. Manja V, Saugstad OD, Lakshminrusimha S. Oxygen Saturation Targets in Preterm Infants and Outcomes at 18-24 Months: A Systematic Review. *Pediatrics*. 2017;139:e20161609.
2. Hwang JS, Rehan VK. Recent Advances in Bronchopulmonary Dysplasia: Pathophysiology, Prevention, and Treatment. *Lung* 2018;196:129-138.
3. Dogra MR, Katoch D, Dogra M. An Update on Retinopathy of Prematurity (ROP). *Indian J Pediatr*. 2017;84:930-936.
4. Cummings JJ, Polin RA; Committee on Fetus and Newborn. Oxygen targeting in extremely low birth weight infants. *Pediatrics* 2016;138:e20161576.
5. Sweet DG, Carnielli V, Greisen G, et al. European consensus guidelines on the management of respiratory distress syndrome – 2019 update. *Neonatology* 2019;115:432-450.
6. Lisa M Askie LM, Darlow BA, Davis PG, Finer N, Stenson B, Vento M, Whyte R. Effects of targeting lower versus higher arterial oxygen saturations on death or disability in preterm infants *Cochrane Database Syst Rev* 2017;4:CD011190
7. Schmidt B, Whyte RK, Shah PS, et al; Canadian Oxygen Trial (COT) Group. Effects of targeting higher or lower oxygen saturations in centers with more versus less separation between median saturations. *J Pediatr* 2016;178:288-291.e2.
8. Tarnow-Mordi W, Stenson B, Kirby A; BOOST-II Australia and United Kingdom Collaborative Groups. Outcomes of two trials of oxygen saturation targets in preterm infants. *N Engl J Med* 2016;374:749-760.
9. Stenson BJ, Tarnow-Mordi WO, Darlow BA, et al. Oxygen saturation and outcomes in preterm infants. *N Engl J Med* 2013;368:2094-2104.
10. Carlo WA, Finer NN, Walsh MC, et al. Target ranges of oxygen saturation in extremely preterm infants. *N Engl J Med* 2010;362:1959-1969.
11. Poets CF, Roberts RS; Schmidt B, et al. Association between intermittent hypoxemia and bradycardia and late death or disability in extremely preterm infants. *JAMA* 2015; 314:595-603.
12. Firme SRE, McEvoy CT, Alconcel C, Tanner J, Durand M. Episodes of hypoxemia during synchronized intermittent mandatory ventilation in ventilator-dependent very low birth weight infants. *Pediatr Pulmonol* 2005;40:9-14.
13. Fathabadi OS, Gale TJ, Dargaville PA. Automated control of inspired oxygen for preterm infants: What we have and what we need. *Biomed Signal Process Control* 2016;28:9-18.
14. Dani C. Automated control of inspired oxygen (FiO₂) in preterm infants: Literature review. *Pediatr Pulmonol* 2019;54:358-363.
15. Hütten MC, Goos TG, Ophelders D, Nikiforou M, Kuypers E, Willems M, Niemarkt HJ, Dankelman J, Andriessen P, Mohns T, Reiss IK, Kramer BW. Fully automated predictive intelligent control of oxygenation (PRICO) in resuscitation and ventilation of preterm lambs. *Pediatr Res* 2015;78:657-63.
16. Dijkman KP, Mohns T, Dieleman JP, van Pul C, Goos TG, Reiss IK, Andriessen P, Niemarkt HJ. Predictive Intelligent Control of Oxygenation (PRICO) in preterm infants on high flow nasal cannula support: a randomised cross-over study. *Arch Dis Child Fetal Neonatal Ed* 2021. Epub ahead of print.
17. Dani C, Pratesi S, Luzzati M, Petrolini C, Montano S, Remaschi G, Coviello C. Cerebral and splanchnic oxygenation during automated control of inspired oxygen (FiO₂) in preterm infants. *Pediatr Pulmonol*. 2021;56:2067-72.
18. Salverda HH, Cramer SJE, Witlox RSGM, Gale TJ, Dargaville PA, Pauws SC, Te Pas AB. Comparison of two devices for automated oxygen control in preterm infants: a randomised crossover trial. *Arch Dis Child Fetal Neonatal Ed* 2021. doi: 10.1136/archdischild-2020-321387.
19. Waitz M, Schmid MB, Fuchs H, Mendler MR, Dreyhaupt J, Hummler HD. Effects of automated adjustment of the inspired oxygen on fluctuations of arterial and regional cerebral tissue oxygenation in preterm infants with frequent desaturations. *J Pediatr* 2015;166:240-244.
20. Gajdos M, Waitz M, Mendler MR, Braun W, Hummler H. Effects of a new device for automated closed loop control of inspired oxygen concentration on fluctuations of arterial and different regional organ tissue oxygen saturations in preterm infants. *Arch Dis Child Fetal Neonatal Ed* 2019;104:F360-F365.
21. Dani C, Pratesi S, Luzzati M, Petrolini C, Montano S, Remaschi G, Coviello C. Cerebral and splanchnic oxygenation during automated control of inspired oxygen (FiO₂) in preterm infants. *Pediatr Pulmonol* 2021;56:2067-2072
22. Poets CF, Franz AR. Automated FiO₂ control: nice to have, or an essential addition to neonatal intensive care? *Arch Dis Child Fetal Neonatal Ed*.2017;102:F5-6.
23. Poets CF, Roberts RS, Schmidt B, Whyte RK, Asztalos EV, Bader D, Bairam A, Moddemann D, Peliowski A, Rabi Y, Solimano A, Nelson H; Canadian oxygen trial investigators. Association between intermittent hypoxemia or bradycardia and late death or disability in extremely preterm infants. *JAMA*. 2015;314:595-603.
24. Salverda HH, Oldenburger NJ, Rijken M, Pauws SC, Dargaville PA, Te Pas AB. The effect of automated oxygen control on clinical outcomes in preterm infants: a pre- and post-implementation cohort study. *Eur J Pediatr*. 2021;180:2107-2113.

GLOBAL HEADQUARTERS

Vyaire Medical, Inc. | 26125 North Riverwoods Blvd | Mettawa, IL 60045, USA

For **International distribution.**

© 2021 Vyaire. Vyaire, the Vyaire logo and all other trademarks or registered trademarks are property of Vyaire Medical, Inc., or one of its affiliates. VYR-GBL-2100241

